IntelliVue MX40

Main screen

Patient demographics

Measurement area

Smart Keys

Local silence

Multi-function button

Alarm area:
Red*/ Yellow*/Inop

Wave 1

Wave 2

Battery status area

Main screen

Main screen

- Touch for 2 seconds to **activate** screen
- Touch again to **change** the screen

*With PIIC Revision N Only

PHILIPS
sense and simplicity

IntelliVue MX40

Smart Keys

Controls available at MX40 that affect the central:

- SpO_2 mode
- Standby
- Relearn

Starts a manual SpO_2 measurement

Initiate a recording at the Information Center

Review or pause alarms or change alarm volume

Activates the Short Range Radio for connection to monitors or Cableless Measurements

Change Mode:

- **Telemetry** – No local alarm volume and screen off at configured time
- **Monitor** – keep screen and volume on at all times

Prints a pre configured report at central

View tabular trends

Change screen ON time

Infinite standby at MX40 and Information Center

To unlock the screen:

1. Touch Smart Keys
2. Touch down arrows
3. Touch the **Unlock** Smart Key

© 2011 Koninklijke Philips Electronics N.V.
All rights are reserved.

Printed in the U.S.A.
4535 642 71291 * SEP 2011

PHILIPS
sense and simplicity